

# Certificate of analysis

---

**Product Name/Description:** DNAiso Reagent  
**Cat. #:** 9770A  
**Lot #:** Specified on product label  
**Storage Conditions:** Room temperature  
**Expiration Date:** Specified on product label or package

**Package Size:** 100 ml

**Package Contents:** 1. DNAiso Reagent 100 ml

**Quality Control Data:**

Genomic DNA of E. coli was extracted using this product according to the user manual. It was confirmed that more than 120 µg of genomic DNA was obtained from 1 ml of bacterial culture of E. coli. When the extracted DNA was dissolved in TE buffer and the absorbance at 230, 260, and 280 nm were measured, the ratio of the absorbance satisfied the following standards.

A260/A280  $\geq$  1.80

A230/A260  $\leq$  1.0

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

**Notice To Purchaser:**

Please refer to the website.

**Product Documents, Patent Markings, Trademarks, & Safety Information:**

Refer to website for availability to download or contact your local office:  
[www.takara-bio.co.jp](http://www.takara-bio.co.jp) (Japanese)  
[www.takarabio.com](http://www.takarabio.com) (English)

**Licence Statements:**

Please visit our website and click the "license" button on the product's web page.

**©2019 Takara Bio Inc. All Rights Reserved.**

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.