



Certificate of Analysis

Product Name/Description: DTI FabSpin Blood DNA Kit II

Cat#: DT9702.25P

Lot#: Specified on product label

Storage conditions: Room temperature

Expiry: Specified on product label

Pack size: 25 Preps

Package contents:

| Component | Quantity |
|------------------------------------|----------|
| Proteinase K (lyophilised) | 1 vial |
| Proteinase K reconstitution buffer | 1 vial |
| Buffer GB | 6 ml |
| Buffer WA | 14 ml |
| Buffer WB | 12 ml |
| Elution buffer | 7 ml |
| PBS | 5 ml |
| Nuclease free water | 1.25 ml |
| Spin columns | 25 pcs |
| Collection tubes | 25 pcs |

Quality Control data:

Yield and Purity test:

The average yield of genomic DNA from fresh collected human blood by using this product was between 3- 5 µg. The ratio of A260/A280 of the genomic DNA solution was from 1.8 to 2.2.

PCR Amplification test:

Expected amplification product was obtained by PCR using genomic DNA extracted from animal whole blood as a template.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of DSS Takara Bio India Pvt Ltd.

This document has been produced electronically and is valid without a signature.